

< XN CHECK™ ASSAY SHEET >

For XN Series

Lot No. **60961102**

Exp. date **2026-06-28**

L2 : Level 2

	RBC 10 ⁶ /μL	HGB		HCT %	MCV fL	MCH		MCHC		PLT 10 ³ /uL
		g/dL	mmol/L			pg	fmol	g/dL	mmol/L	
TARGET	4.20	11.5	7.1	33.9	80.6	27.4	1.699	33.9	21.1	221
LIMIT	0.21	0.3	0.2	3.4	4.0	2.2	0.136	4.7	3.0	33

	RDW-SD fL	RDW-CV %	PDW fL	PCT %	MPV fL	P-LCR %
TARGET	45.0	15.7	8.3	0.20	9.2	14.1
LIMIT	4.5	1.6	1.3	0.06	0.8	21.2

	WBC 10 ³ /μL	WBC-D 10 ³ /μL	WBC-P 10 ³ /μL	NEUT# 10 ³ /μL	LYMPH# 10 ³ /μL	MONO# 10 ³ /μL	EO# 10 ³ /μL	BASO# 10 ³ /μL	IG# 10 ³ /μL	NRBC# 10 ³ /μL	HPC# 10 ³ /μL	AS-LYMP# 10 ³ /μL
TARGET	6.98	7.13	7.24	3.09	2.03	0.78	0.75	0.33	0.78	0.43	0.072	0.33
LIMIT	0.42	0.43	0.65	0.46	0.41	0.47	0.38	0.26	0.23	0.22	0.029	0.17

	NEUT% %	LYMPH% %	MONO% %	EO% %	BASO% %	IG% %	NRBC% %	HPC% %	AS-LYMP% %	NEUT-RI FI	NEUT-GI SI
TARGET	44.4	29.0	11.1	10.7	4.8	11.2	6.2	0.99	4.8	88.9	171.3
LIMIT	6.7	5.8	6.7	5.4	3.7	3.4	3.1	0.40	2.4	35.6	17.1

	PLT-F 10 ³ /μL	IPF# 10 ³ /μL	IPF %	RE-LYMP# 10 ³ /μL	RE-LYMP% %
TARGET	227	44.9	19.8	0.52	7.4
LIMIT	68	33.7	5.0	0.26	3.7

	RET# 10 ⁶ /μL	RET% %	RBC-O 10 ⁶ /μL	HFR %	MFR %	LFR %	IRF %	RET-He		RBC-He		PLT-O 10 ³ /μL
								pg	fmol	pg	fmol	
TARGET	0.0990	2.36	4.12	3.0	22.0	75.0	25.0	24.2	1.502	28.2	1.750	224
LIMIT	0.0297	0.47	0.41	11.0	28.0	33.0	36.0	2.4	0.150	5.6	0.350	67

	HYPO-He %	HYPER-He %	MicroR %	MacroR %
TARGET	0.6	0.3	7.3	3.9
LIMIT	4.0	0.7	5.1	1.6

There are some parameters which is not displayed on IPU according to the instrument.

Please refer to the package insert for the handling of the XN CHECK.

Do not leave control blood in the room temperature over an hour.

Please store it in a refrigerator (2-8°C) immediately after use.