

< XN CHECK™ ASSAY SHEET >

For XN Series

Lot No. **60961103**

Exp. date **2026-06-28**

L3 : Level 3

	RBC 10 ⁶ /μL	HGB		HCT %	MCV fL	MCH		MCHC		PLT 10 ³ /uL
		g/dL	mmol/L			pg	fmol	g/dL	mmol/L	
TARGET	4.91	14.6	9.1	42.2	86.0	29.7	1.845	34.6	21.5	497
LIMIT	0.25	0.4	0.3	4.2	4.3	2.4	0.148	4.8	3.0	45

	RDW-SD fL	RDW-CV %	PDW fL	PCT %	MPV fL	P-LCR %
TARGET	48.6	15.8	8.5	0.45	9.1	13.4
LIMIT	4.9	1.6	1.0	0.11	0.6	20.1

	WBC 10 ³ /μL	WBC-D 10 ³ /μL	WBC-P 10 ³ /μL	NEUT# 10 ³ /μL	LYMPH# 10 ³ /μL	MONO# 10 ³ /μL	EO# 10 ³ /μL	BASO# 10 ³ /μL	IG# 10 ³ /μL	NRBC# 10 ³ /μL	HPC# 10 ³ /μL	AS-LYMP# 10 ³ /μL
TARGET	16.20	16.46	16.75	7.84	3.70	2.08	1.80	0.78	1.96	1.11	0.170	0.61
LIMIT	0.97	0.99	1.01	1.18	0.74	1.04	0.90	0.61	0.49	0.56	0.051	0.31

	NEUT% %	LYMPH% %	MONO% %	EO% %	BASO% %	IG% %	NRBC% %	HPC% %	AS-LYMP% %	NEUT-RI FI	NEUT-GI SI
TARGET	48.4	22.8	12.9	11.1	4.8	12.1	6.9	1.02	3.7	82.4	171.4
LIMIT	7.3	4.6	6.5	5.6	3.7	3.0	3.5	0.31	1.9	33.0	17.1

	PLT-F 10 ³ /μL	IPF# 10 ³ /μL	IPF %	RE-LYMP# 10 ³ /μL	RE-LYMP% %
TARGET	485	96.5	19.9	0.94	5.8
LIMIT	73	67.6	5.0	0.47	2.9

	RET# 10 ⁶ /μL	RET% %	RBC-O 10 ⁶ /μL	HFR %	MFR %	LFR %	IRF %	RET-He		RBC-He		PLT-O 10 ³ /μL
								pg	fmol	pg	fmol	
TARGET	0.0626	1.27	4.83	2.6	19.8	77.6	22.4	25.4	1.576	29.9	1.856	491
LIMIT	0.0188	0.44	0.48	11.0	28.0	33.0	36.0	2.5	0.158	6.0	0.371	74

	HYPO-He %	HYPER-He %	MicroR %	MacroR %
TARGET	0.3	0.5	3.8	4.4
LIMIT	3.0	1.0	3.0	1.8

There are some parameters which is not displayed on IPU according to the instrument.

Please refer to the package insert for the handling of the XN CHECK.

Do not leave control blood in the room temperature over an hour.

Please store it in a refrigerator (2-8°C) immediately after use.